

SLOVENSKÁ NÁRODNÁ AKREDITAČNÁ SLUŽBA

Karloveská 63, 840 00 Bratislava 4, Slovenská republika

CERTIFICATEOF ACCREDITATION

No. K-151

The Slovak National Accreditation Service based on the decision No. 758/9250/2021/1 dated 29.06.2021 certifies that

VYPRO, s.r.o. Calibration laboratory VYPRO

K zábraniu 1653, 911 01 Trenčín ID Number: 36 705 942

is competent to perform calibration of measuring instruments of force and force sensors within the accreditation scope delineated in the Annex to this Certificate. The Annex is an integral part of Certificate of Accreditation.

The accredited body gives evidence of competence to perform the accredited activity impartially and trustworthily by meeting the requirements of the ISO/IEC 17025: 2017 Standard.

Accreditation granted on 29.06.2021 is valid until 29.06.2026.

Bratislava 29.06.2021

SNAS SNAS

Martin Senčák

The Annex is an integral part of the Certificate of Accreditation

Accreditation scope

Name of the accredited body:

VYPRO, s.r.o.

Calibration laboratory VYPRO K zábraniu 1653, 911 01 Trenčín

Laboratory with fix accreditation scope.

Item	Kind of measuring instrument/ measurement means	Measurement range	Expanded uncertainty $U(k=2)$	Established Methods		
				Kind/ Principle	Identification	Other specifications
1	Measuring instruments of force and force sensors	(0.1 - 2) kN	0.1 % RD ¹⁾ tension 0.1 % RD ¹⁾ compression	Direct comparison with standard	ČSN EN ISO 376 (PP-Force gauges)	In laboratory
		(2 - 200) kN	0.07 % RD ¹⁾ tension 0.07 % RD ¹⁾ compression			

Notes:

1) RD – Reading value

